

1 Frequently Asked Questions

1.1 `align_mQTL()` aborted the processing with the error message "Peak validation threshold exceeds spectrum maximum and minimum value"

The mQTL.NMR package suggests default values optimized through extensive experimental setup and during several NMR studies for alignment using the RSPA approach. However, these parameters might not be well adapted for the data under consideration due to several factors (intrinsic data structure, inappropriate normalisation/scaling). For this reason we have made available the function `setupRSPA()` enabling a simple modification of different parameters. In particular, the error message "Peak validation threshold exceeds spectrum maximum and minimum value" indicates that the peak amplitude threshold parameter (`ampThr`) defined inside the `peakParam` structure is not well optimized for the considered data. In such a case we advise the user to try one of the following solutions:

I- Try a different normalization/scaling method: in some cases normalization/scaling step affect drastically the structure of data and more adapted methods should be selected.

II- Modify the alignment parameters through the function `setupRSPA()`:

In this specific case the user have two options:

- Manual: by a careful optimization of alignment parameters with the respect to the characteristics of his data.
- Automated: but less optimal approach concerns the use of the function `getAmpThr()` available already in the mQTL.NMR package (and accessible by `mQTL.NMR::getAmpThr`) which allows to estimate automatically this parameter. In this specific case, the user needs only to discard the assignment instruction (of `PeakParam$ampThr`) from the list of parameters defined inside the function `setupRSPA()`. The estimation process of the parameter will be then launched automatically upon call.

In order to perform one of the two options mentioned in the solution (II), the user needs first to retrieve the original function `setupRSPA()` in his workspace and rename it in order to modify the values of parameters accordingly. Once the parameter modification finished the original function should be exchanged by the modified one (e.g. `MysetupRSPA`) in the package namespace by using the following script:

```
> ## Exchange the function setupRSPA() in the namespace of the mQTL.NMR package
> unlockBinding("setupRSPA", as.environment("package:mQTL.NMR"))
> assignInNamespace("setupRSPA", MysetupRSPA, ns="mQTL.NMR",
+   envir=as.environment("package:mQTL.NMR"))
> assign("setupRSPA", MysetupRSPA, as.environment("package:mQTL.NMR"))
> lockBinding("setupRSPA", as.environment("package:mQTL.NMR"))
> ppmToPt<-mQTL.NMR::ppmToPt
```

1.2 I want to use the package only for the preprocessing of my NMR data and do not have neither genomic nor clinical data

Only some macro functions require the simultaneous use of metabolomic and genomic data. The user can perform the preprocessing steps using the micro functions provided also by the mQTL.NMR package (e.g. for normalization use simply the function `normalise()`) which handles only R objects (data frames and matrices).

1.3 `align_mQTL()` aborted the processing with the error message "Error in if (step >= splength)...missing value where TRUE/FALSE needed"

The error is likely due to the inappropriate format of the input files. Please make sure that the adopted format is supported by the mQTL.NMR package.

2 Session Information

The version number of R and packages loaded for generating the vignette were:

R version 3.2.0 RC (2015-04-08 r68161)

Platform: x86_64-apple-darwin10.8.0 (64-bit)

Running under: OS X 10.6.8 (Snow Leopard)

locale:

[1] en_US.UTF-8/en_US.UTF-8/en_US.UTF-8/C/en_US.UTF-8/en_US.UTF-8

attached base packages:

[1] stats graphics grDevices utils datasets methods base

loaded via a namespace (and not attached):

[1] BiocStyle_1.6.0 tools_3.2.0